



Algoritmo BLS-DAE-SRC 2015 secondo ILCOR CoSTR*

Rapido controllo concernente la sicurezza
della scena
(soprattutto intossicazioni ed elettricità)

Persona priva di coscienza
Respirazione mancante o
anormale

Chiamare aiuto ad alta voce
Allarmare Tel 144
Prendere o richiedere il DAE

30 compressioni toraciche (100 – 120 /min)
seguite da 2 ventilazioni
oppure
compressioni toraciche senza ventilazioni

DAE sul luogo
Seguire le istruzioni dell'apparecchio

Prestare attenzione a:

- Compressioni toraciche **5 – 6 cm** risp. **1/3** del diametro della cassa toracica nei bambini
- Interruzione **minimale**
- Luogo di compressione **al centro** della cassa toracica
- **Rilascio** completo
- Ventilazioni: movimenti **visibili** del torace!

Origini:

- Travers AH, et al., 2015 *International Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care Science With Treatment Recommendations. Part 3: adult basic life support and automated external defibrillation: Circulation. 2015;132 (suppl 1):S51–S83.
- Maconochie IK, et al., 2015 International Consensus on Cardiopulmonary Resuscitation and Emergency Cardiovascular Care Science with Treatment Recommendations Part 6: Pediatric basic life support and pediatric advanced life support., Resuscitation 95 (2015) e147–e168